

TM 027 – ANTIFUNGAL ASSAY AGAR

INTENDED USE

For assay of antifungal activity in pharmaceutical and other products by cylinder plate/disc method.

PRODUCT SUMMARY AND EXPLANATION

Fungal infections have been reported to have dramatically increased in the past decade, and these often occur as systemic infections or as co-infections with other diseases, such as AIDS or cancer, or in patients who are immunocompromised. Unfortunately, in addition to the limited number of antifungal drugs currently available, fungal infections tend to rapidly develop resistance to these drugs. For these reasons, fungal infections now show much higher mortality rates than bacterial infections. The rapid increase in fungal infections and the growing number of new antifungal agents indicate an increasing need for rapid and accurate methods for antifungal screening and susceptibility testing. Antifungal Assay Agar was formulated by Berger and Lazecka for convenience in assaying antifungal activity of pharmaceutical products and other materials by both base and seed layers for assays by cylinder plate or disc methods.

Assay Methods Cylinder plate method: This method was first devised by Abraham et al and later modified by Schmidt and Moyer and it depends upon diffusion of the antibiotic from vertical steel cylinders placed on the surface of inoculated agar medium. This produces zones of inhibition around the cylinder containing antibiotic solution depending upon the concentration of the antibiotic in the cylinder. This method is commonly employed in the assay of pharmaceutical preparations of Penicillin and other antibiotics. For assay, use Petri plates with 20 x100 mm dimension and stainless steel or porcelain cylinders with the outside diameter 8 mm, inside diameter 6 mm and length 10 mm. All dimensions should have a tolerance of 0.1 mm. The cylinders should be carefully cleaned to remove all the impurities. For assays requiring base and seed layer, the base layer is allowed to solidify first and then overlaid with the seed agar containing the proper concentration of the test organism. Most assays require base layer of 21 ml and seed layer of 4 ml. Generally, 6 cylinders are used per plate. The cylinders are placed on inoculated plates at equal distance.

Paper-disc method: Paper discs with a diameter of 9 mm are impregnated with the antibiotic solution and placed on the culture medium. Antibiotic can also be applied to the disc after it has been placed on the medium. Plates containing a single layer of medium with 2 mm thickness may be used for these tests. All other steps are similar to the cylinder plate method.

COMPOSITION

Ingredients	Gms / Ltr
Dextrose (Glucose)	50.000
Sodium citrate	4.500
Potassium dihydrogen phosphate	0.550
Citric acid	1.000
Tryptone	4.000
Pyridoxine hydrochloride	0.00025
Thiamine	0.00025
Inositol	0.025
Calcium pantothenate	0.0025
Niacin	0.0025
Potassium chloride	0.425
Calcium chloride	0.125
Magnesium sulphate	0.125
Ferric chloride	0.0025

Manganese sulphate	0.0025
Biotin	0.000008
Agar	15.000

PRINCIPLE

The defined ingredients in the medium provide the necessary nutrients and growth factors required for the development of the test culture. Phosphate is included in this medium for good buffering action. Dextrose in the medium serves as a carbon and energy source. Other ingredients like the sulphates; vitamins, growth factors etc are added to enhance the growth of the test organisms, so that the inhibition obtained is always due to the antifungal agents and not due to nutrient depletion.

INSTRUCTION FOR USE

- Dissolve 75.76 grams in 1000 ml purified / distilled water.
- Heat to boiling to dissolve the medium completely.
- Sterilize by autoclaving at 15 psi pressure (121°C) for 15 minutes. Cool to 45-50°C.
- Mix well and pour into sterile Petri plates.

QUALITY CONTROL SPECIFICATIONS

- Appearance of Powder** : Cream to beige homogeneous free flowing powder.
- Appearance of prepared medium** : Light yellow coloured clear to slightly opalescent gel forms in Petri plates.
- pH (at 25°C)** : 5.5±0.2

INTERPRETATION

Cultural characteristics observed after incubation.

Microorganism	ATCC	Inoculum (CFU/ml)	Growth	Recovery	Incubation Temperature	Incubation Period
<i>Saccharomyces cerevisiae</i>	9763	50-100	Luxuriant	>=70%	25-30°C	18-48 Hours
<i>Aspergillus brasiliensis</i>	16404	50-100	Luxuriant	>=70%	25-30°C	18-48 Hours

PACKAGING:

In pack size of 500 gm bottles.

STORAGE

Dehydrated powder, hygroscopic in nature, store in a dry place, in tightly-sealed containers between 25-30°C and protect from direct sunlight. Under optimal conditions, the medium has a shelf life of 4 years. When the container is opened for the first time, note the time and date on the label space provided on the container. After the desired amount of medium has been taken out replace the cap tightly to protect from hydration.

Product Deterioration: Do not use if they show evidence of microbial contamination, discoloration, drying or any other signs of deterioration.













DISPOSAL



After use, prepared plates, specimen/sample containers and other contaminated materials must be sterilized before discarding.

REFERENCES

1. Abraham, Chain, Fletcher, Florey, Gardner, Heatley and Jennings, 1941, Lancet ii: 177.
2. Beck-Sague C. and Jarvis W. R., 1993, J. Infect. Dis., 167:1247-1251.
3. Berrouane Y. F., Herwaldt L. A., and Pfaller M. A., 1999, J. Clin. Microbiol., 37:531-537.
4. Isenberg, H.D. Clinical Microbiology Procedures Handbook 2nd Edition.
5. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock, D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
6. Schmidt and Moyer, 1944, J. Bacteriol., 47:199. 7. Weinstein, M. P., Towns M. L., Quartey S. M., Mirrett S., Reimer L. G., Parmigiani G. and Reller L. B., 1997., Clin. Infect. Dis., 24:584-602.

 GMP Good Manufacturing Practices Certified	 IVD For In Vitro Diagnostic Use	 QTY. Quantity	 LOT/ B. NO. Lot / Batch Number	 REF Catalogue Number	 Manufacturer
 Temperature Unit	 EC REP Authorized Representative <small>MedNet GmbH Borkstrasse 10, 48163 Muenster, Germany</small>	 European Conformity	 QR Code	 Consults Instructions for Use	 Best Before

NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

***For Lab Use Only**
Revision: 08 Nov., 2019