

## TS 277 – FRIIS SUPPLEMENT

### INTENDED USE

A selective supplement for detection of non-avian Mycoplasmas in pharmaceutical products in accordance with European pharmacopoeia.

### COMPOSITION

Ingredients	Concentration
Bacitracin	125.0mg/vl
Methicillin	125.0mg/vl
Nalidixic acid	10.0mg/vl

(per vial sufficient for 1000 ml medium)

### Applied in Medium













TM 2095 - FRIIS LIQUID MEDIUM BASE

### INSTRUCTION FOR USE

Rehydrate the contents of 1 vial aseptically with 2 ml of 0.2N NaOH and 8 ml of sterile distilled water. Mix well and aseptically add to 1000 ml of sterile, cooled (45-50°C) TM 2095 - FRIIS LIQUID MEDIUM BASE. Mix well and dispense as desired.

### STORAGE

Vials should be stored in sealed container at 2-8°C.

 GMP Good Manufacturing Practices Certified	 IVD For In Vitro Diagnostic Use	 QTY. Quantity	 LOT/ B. NO. Lot / Batch Number	 REF Catalogue Number	 Manufacturer
 Temperature Unit	 EC REP Authorized Representative <small>MaxMer GmbH Buckhorn 16, 48147 Münster, Germany</small>	 European Conformity	 QR Code	 Consults Instructions for Use	 Best Before

**NOTE:** Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

**\*For Lab Use Only**  
Revision: 28 Feb., 2022