

TS 295 – CHROMOGENIC SUPPLEMENT

INTENDED USE

For the enumeration of faecal coliform by membrane filter technique.

COMPOSITION

Ingredients	Concentration
Chromogenic Substrate	0.100g

(Per vial sufficient for 1000 ml medium)

Applied in Medium













TM 2186 - M-FC BASAL MEDIUM

INSTRUCTION FOR USE

Rehydrate the contents of one vial aseptically with 5 ml sterile distilled water. Mix well and aseptically add it to 1000 ml of sterile, molten, cooled (45-50°C) TM 2186 - M-FC BASAL MEDIUM. Mix well and pour into sterile Petri plates.

STORAGE

Vials should be stored in sealed container at 2-8°C.

 GMP Good Manufacturing Practices Certified	 IVD For In Vitro Diagnostic Use	 QTY. Quantity	 LOT/ B. NO. Lot / Batch Number	 REF Catalogue Number	 Manufacturer
 Temperature Unit	 EC REP Authorized Representative <small>MaxMer GmbH Buckhorn 10 48163 Ahaus, Germany</small>	 European Conformity	 QR Code	 Consults Instructions for Use	 Best Before

NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

***For Lab Use Only**
Revision: 8 March., 2022